


www.balteau.com



The generator is driven by an electronic circuit board (the power components are cooled by forced air). Components can be readily installed in the crawler body and connected to its general circuitry. Any DC voltages within the range of 72 to 130v can be selected and applied to the generator.

An exclusive feature allows generator to operate at constant kV and mA no matter of the battery cycles. This feature ensures a stable dose output and guarantees image quality consistency all along the exposure time.

This specific process, is also improving the operating possibilities of the battery pack. Indeed much more exposures can be taken. This process improves the battery lifetime thanks to a better management of discharge cycles.



Specifications	Units	BALTOSPOT GFC160C
High voltage range:	kV	112 - 160
Approx. tube output	mA	3
Beam angle:		360° x 40°
Focal spot size (IEC336):		4 x 0,9
Max. current:	A	4,5
Steel penetration at max (*):	mm	12,5
Working temperature:	°C	- 10 to + 40
Inherent filtration:	mm	7 AI
Anode cooling type:		Air
Duty cycle:	%	50% at 35°C
Insulation:		SF6 Gas
Mains:	V	72 - 130
Weight (without rings):	Кд	11
Tubehead:		Glass
Control unit:		CF3

^{*} FFD= 700mm/Film Carestream Industrex AA400/D=2.0/t=10min.



Russell Fraser Sales Pty Ltd

Unit 7/38 Waratah St Kirrawee NSW 2232 Ph: (02) 9545 4433 Fax: (02) 9545 4218

Email: rfs@rfsales.com.au Web: www.rfsales.com.au